

RESULT OF ANALYSIS

Name of the Material / Sample	Polythene covers	Test Method Reference	In-House
Manufacturer Name	Lotus Polymers Industries	Date of Analysis	11/03/2022
In house AR No.	PLS/22/TR/0074		
Name and address of customer	Lotus Polymers Industries, B-78 & B-79, KSSIDC Ind. Estate, Bommasandra, Hosur Road, Bangalore-560099		

Name of Test	Specification	Result
Bioburden test for Polythene Covers by membrane filtration technique	NMT 50 CFU / 3 covers	02 CFU/ 3 Covers

Remarks: Nil



Prepared by	Reviewed by	Approved by (QA)
<i>[Signature]</i> 19/03/2022	<i>[Signature]</i> 19/03/22	<i>[Signature]</i> 19/03/2022

COMPANY CONFIDENTIAL